

Problem: questionable EC readings

Note: occasional negative values are not necessarily a significant problem, provided they are intermittent and less than 5% of the data. If field conditions are rough or dry, the coulter-electrodes may not be making consistent contact with moist soil, resulting in negative values.

